



*New Brunswick Laboratory*  
*U.S. Department of Energy*

# Certificate of Analysis

## CRM U900

### Uranium Isotopic Standard

10 mg Uranium as  $U_3O_8$

	$^{234}U$	$^{235}U$	$^{236}U$	$^{238}U$
<b>Atom Percent:</b>	<b>0.7777</b>	<b>90.196</b>	<b>0.3327</b>	<b>8.693</b>
<b>Uncertainty:</b>	<b>±0.0015</b>	<b>±0.011</b>	<b>±0.0010</b>	<b>±0.008</b>
<b>Weight Percent:</b>	<b>0.7735</b>	<b>90.098</b>	<b>0.3337</b>	<b>8.795</b>

This Certified Reference Material (CRM) is primarily intended for the calibration of mass spectrometers used to perform uranium isotopic measurements. The specific purpose of this isotopic standard is for the determination of mass discrimination effects for uranium isotopes being measured under similar analytical conditions. Each unit of CRM U900 consists of approximately 10 milligrams of uranium, in the form of highly purified  $U_3O_8$ , contained in a glass bottle.

The indicated uncertainties for the isotopic composition of the CRM are 95% confidence intervals for a single determination. This term can be defined as an approximate two-sigma limit, where sigma is the standard deviation of the measurements data obtained from the material. The uncertainties include allowances for inhomogeneity of the material as well as analytical error.

This CRM was originally issued in 1970 by the National Bureau of Standards (NBS) as Standard Reference Material (SRM) U-900. The measurements made at NBS leading to the certification were performed by E. L. Garner, L. A. Machlan, M.S. Richmond and W. R. Shields. In 1987, the technical and administrative transfer of NBS Special Nuclear SRMs into the NBL CRM Program was coordinated by the NBS Office of Standard Reference Materials and N. M. Trahey, NBL.

The certified isotopic abundance values were determined using a solid-sample thermal ionization mass spectrometer equipped with a Faraday cup detection system. The measured  $^{238}U$  values were calculated from the  $^{235}U/^{238}U$  values, which were corrected for mass discrimination effects by intercomparison with synthetic calibration mixtures of similar  $^{235}U$  levels, prepared from high-purity  $^{235}U$  and  $^{238}U$  separated isotopes. The  $^{235}U/^{238}U$  value for this standard, 10.375, is known to at least 0.1%.

The  $^{234}U$  and  $^{236}U$  abundances were determined by isotope dilution mass spectrometry using high-purity  $^{233}U$  as the spike.

NOTE: NBS Special Publication 260-27 presents further details of the measurements made at NBS which provided the basis for the certification, and is available from the NBS Office of Standard Reference Materials upon request.

March 30, 2008  
Argonne, Illinois

[www.nbl.doe.gov](http://www.nbl.doe.gov)  
Page 1 of 1

Jon Neuhoff, Director  
New Brunswick Laboratory

(Editorial revision of Certificate dated October 1, 1987)